

TECHNICAL INFORMATION

OPTIMA GROUND



LEG-FOOT DEVICE

Pre- and post-operative treatment of lower limbs, compound fractures, rehabilitation, stabilisation.

PRODUCT COMPONENTS

- 1- Leg
- 2- Lining
- 3- Single
- 4- Bottom
- 5- Velcro



The Optima line is indicated for the treatment of alterations in the morphology of the foot.

Optima Molliter S.r.l. declares that the Optima tutors are medical devices conceived and designed according to the state of the art, have been checked and made compliant with EU Regulation 2017/745 for the applicable parts. The braces comply with IWGDF's international guidelines for the management of diabetic foot ulcers. <https://iwgdfguidelines.org/guidelines/guidelines>.

THERAPEUTIC INDICATIONS

- Plantar diabetic ulcer with vascular complications in the leg, neuroischaemic ulcer before and/or after revascularisation (using the OPTIMA PUZZLE KIT 3x3 in addition).
- Management of the foot contralateral to the injured foot or management of the foot at risk of ulceration.
- Compound fracture of the metatarsals.
- Preoperative stabilisation.

TECHNICAL FEATURES

Predisposed components directly adaptable to the patient, heat- mouldable high stiffener, breathable and washable upper.

Adjustable structure with velcro closures. Wide openings to allow inspection and ventilation. Rigid anti-flexion and anti-torsion midsole, outsole with movement control and redistribution of the pressure load.

OPTIMA GROUND does not need any specific contralateral compensation footwear, but it is compatible with any model of the YDA and MAC lines.

TECHNICAL CHARACTERISTICS

MATERIALIALS:

UPPER: PU 70% + PES 20% + COT 10%

LINING POLYESTER: 100%

INSOLE INSERT: POLYAMIDE PA6 70% +GLASS FIBER 30%

BOTTOM SOLE: POLYURETHANE

PROTECTIVE SOCK: 90% COTTON, 7% POLYAMIDE, 3% ELASTANE

OPTIMA PRODUCTS ARE LATEX FREE.

CLOSURE: 3 VELCROS

PACKAGING: WEIGHT: GR 205 / MATERIAL: CARTON

OPERATIVE TEMPERATURE AND STORAGE: - 5° | + 45°

PRODUCTION LOT: AS DESCRIBED ON THE LABEL ON THE FRONT OF THE BOX

BASE COMPONENTS: MONOLAYER

ADDITIVES: MONOLAYER / OPTIMA KIT 3X3 PLANTAR INSOLE / PROTECTIVE SOCK

PRODUCT CLEANING


To clean the product, use warm or cold water and neutral soap; rinse once the operation is completed. Avoid any corrosive liquids such as alcohol, laundry detergents, bleach, ammonia, and solvents. Avoid using abrasive materials such as brushes, steel wool, and sand paper, or a knife blade to remove stains. Optima Molliter S.r.l. shall NOT be held liable for damages caused by abrasive products, by products which attack the materials, by direct exposure to heat sources.

Class I medical device according to Regulation (EU) 2017/745

In accordance with Recommendations 2-5-6-7c-9 of the IWGDF International Guidelines on the Management of the Diabetic Foot. <https://iwgdfguidelines.org/guidelines/guidelines>

TAGLIE

	S 34-36	M 37-39	L 40-42	XL 43-45	XXL 46-48
A MM.	247	267	288	312	338
B MM.	106	115	120	128	138
C MM.	222	226	230	240	254
PESO GR. senza plantare	325	378	480	518	604



Optima Molliter Srl

Company Certified under: ISO 9001:2015 IT346378 - ISO 13485:2016 IT346377

Via Breda 19/21 62012 Civitanova Marche (MC) Italy T +39 0733 801060

info@optimamolliter.com / www.yda.it / www.optimamolliter.com

